

GEM Sample Line Leak Test

Install a 5 ltr / min flow meter inline with the Nafion sample line, as shown below Figure 1.

1. In PlusRA, from the Control Menu, select the Set GEM COCH4.
2. Check the sample line for leaks by using your finger to press down on the needle to occlude the sample port in the Gem valve, As in Figure 2.
3. In PlusRA, from the Control Menu, select SetALLOFF (CPL / GEM).




Figure 2.




Figure 1.

